

IV-E220A To Determine The Conductivity Of Liquid By Kohlrausch Method



Specifications:

- A circular slide wire is calibrated from 0.01 to 10 ohms
- A rotary dial furnishes six multipliers of 0.1, 1, 10, 100 & 1000 & 10000.
- Range of measurement is 0.01 to 100 K Ohm
- Terminals for supply and head phone/detectors are provided on the panel.
- A detector key is also provided.
- The position of slider is adjusted to the position of minimum sound or position of minimum A.C. signals.
- Accuracy of coils $\pm 0.1\%$
- Accuracy of measurement $\pm 1\%$ at mid scale & $\pm 2\%$ overall.

Optional Accessories:

- Kohlrausch Conductivity Bridge
- Induction coil
- Conductivity Probe With Beaker
- Battery Eliminator 2-12V, 2A
- Head Phone/ Galvanometer/Digital Multimeter.