IVESTY

GURU KRIPA ENTERPRISES

AN ISO 9001:2015 & NSIC CERTIFIED COMPANY

Sur No - 668, Plot No - 268 , Fatevadi, Sarkhej , Ahmedabad,Gujarat - 382210

Email: info@gkelabs.com,www.gkelabs.com,Mobile::9106316361

GEM PORTAL REGISTERED

IV-E650 To Find The Coefficient Of Mutual Inductance Of Two Coil Using Rayleigh's Method

Scope of Learning:

- TO FIND THE COEFF. OF MUTUAL INDUCTANCE OF TWO COILS USING RAYLEIGH'S METHOD

Technical Specification:

- Power Supply: 0 - 2V DC, at 500mA
- Digital Meter: 20V DC
- Variable Resistor: 1K
- Key: 4 Way Key
- Key: One Push Key
- Resistor: 3 Resistors
- Ext. Rheostat: Rheostat (if Required)
- On/Off Key: On-Off Key
- Mains: 230V AC ±10%, 50Hz
- Fuse: 500Ma
- Dimension of Unit (mm): W 215 x D 195 x H 130

Salient Features:

- Front panel built with high class insulated Printed Circuit Board sheet with well printed circuits and symbols.
- Ballastic Galvanometer With Lamp and Scale Arrangement.
- Instruction manual.
- Connections are brought out through 4mm Colored Sockets.
- Patch Cords 4mm.
- The trainer is housed in ABS Plastic cabinet.
- Size of the trainer set 12"x10"

Optional Accessories

- No